

# Progress of Product Registration Working Group

4<sup>th</sup> Joint Conference of Taiwan and  
Japan on Medical Products Regulation



**Ta-Jen Wu**

Product Registration WG

2016.12.07

**衛生福利部食品藥物管理署**

Food and Drug Administration,  
Ministry of Health and Welfare

<http://www.fda.gov.tw/>

# Overview

- 1. Product Registration(PR) WG**
2. Regulations on Product Pre-market Registration in Taiwan
3. Regulations on Product Pre-market Registration in Japan
4. Review Requirements of Dental Implants in Taiwan and Japan
5. Case Study
6. Summary



# Members of PR Working Group

Taiwan	
Ms. Ting-Yao Wu	Section Chief, Division of Medical Devices & Cosmetics, Food and Drug Administration, Ministry of Health and Welfare
Dr. Ta-Jen Wu	Technical Specialist, Division of Medical Devices & Cosmetics, Food and Drug Administration, Ministry of Health and Welfare
Dr. Han-Son Dawn	Regulation Director, Taiwan Medical and Biotech Industry Association
Mr. Tzu-Wei Li	Industrial Technology Research Institute
Japan	
Dr. Madoka Murakami	Office of International Programs Pharmaceuticals and Medical Devices Agency (PMDA)
Mr. Hirokazu Takahashi	Regulatory System Committee, Japan Federation of Medical Devices Associations (JFMDA)
Mr. Makoto Yokote	Asia Subcommittee, Japan Federation of Medical Devices Associations (JFMDA)

# Product Registration Working Group

藥求安全 食在安心

- **Goal:**

1. Share the reviewing experiences between Taiwan and Japan.
2. Mutual understanding
3. Simplified registration processes to benefit both industries.

- **Working item in 2016:**

Implement a case study program of dental implant to understand the review differences between Taiwan (TFDA) and Japan (PMDA).

# Overview

1. Product Registration(PR) WG
- 2. Regulations on Product Pre-market Registration in Taiwan**
3. Regulations on Product Pre-market Registration in Japan
4. Review Requirements of Dental Implants in Taiwan and Japan
5. Case Study
6. Summary



# Regulations on Product Pre-market Registration in Taiwan

藥求安全 食在安心

- **Pharmaceutical Affairs Act**
  - For the manufacturing and import of medical devices, an application together with fees paid, shall be filed with the central competent health authority for registration and market approval.
- **Regulations for Governing the Management of Medical Device**
  - Medical device classification
- **Regulation for Registration of Medical Device**
  - Registration requirements



# Risk Based Regulation

藥求安全

食在安心

Low risk

High risk

GMP/QSD

GMP/QSD

GMP/QSD

Class 1

Class 2

Class 3

affidavit

On-site  
registration

Technical  
Document

Technical  
Document

*QSD: Quality system  
documentation*


Preclinical testing,  
QC documents and  
Clinical trial reports  
(conditional).

Documents  
required for  
Registration

■ More than 1700 regulated medical device items in Taiwan

# Before Marketing a Medical Device Product in Taiwan

藥求安全 食在安心









# Overview

1. Product Registration(PR) WG
2. Regulations on Product Pre-market Registration in Taiwan
- 3. Regulations on Product Pre-market Registration in Japan**
4. Review Requirements of Dental Implants in Taiwan and Japan
5. Case Study
6. Summary




# Risk Based Regulation

藥求安全 食在安心

GHTF Classification		PAL classification		
		As of April, 2015		
		Category	Regulatory requirements	Japanese MD Nomenclature
 <b>Class A</b> <b>Extremely low risk</b> e.g., X-ray film	<b>General MDs (Class I)</b>	<b>Self declaration</b> Approval of the product is not required, but marketing notification is necessary.	<b>1,195</b>	
 <b>Class B</b> <b>Low risk</b> e.g., MRI, digestive catheters	<b>Controlled MDs (class II)</b>	<b>Third party Certification</b> Certification by a registered certification body is required. • Certification standard	<b>1,955</b> <b>(1,369 for 3<sup>rd</sup> Party)</b>	
 <b>Class C</b> <b>Medium risk</b> e.g., dialyzer	<b>Specially Controlled MDs (class III &amp; IV)</b>	<b>Certification (Review by CB)</b>	<b>Minister's Approval (Review by PMDA)</b> The Minister's approval for the product is required. • Approval standard • Review guideline	<b>760</b>
 <b>Class D</b> <b>High risk</b> e.g., pacemaker		<b>346</b>		

# Overview of review process

藥求安全 食在安心



# Overview

1. Product Registration(PR) WG
2. Regulations on Product Pre-market Registration in Taiwan
3. Regulations on Product Pre-market Registration in Japan
- 4. Review Requirements of Dental Implants in Taiwan and Japan**
5. Case Study
6. Summary

# Requirements of Dental Implants

藥求安全 食在安心

Items	Taiwan	Japan
1. Classification	Class III	Class III
2. Application Documents	Almost the same	
3. STED documents	Almost the same	



Items	Taiwan	Japan
4. Administrative Information	<ul style="list-style-type: none"> <li>(1) Pharmaceutical Company License or Certificate</li> <li>(2) Affidavit A</li> <li>(3) Certificate to Foreign Government (only for foreign manufacturers)</li> <li>(4) Letter of Authorization (only for foreign manufacturers)</li> <li>(5) Certificate of GMP for domestic or QSD for foreign manufacturers</li> </ul>	<p>Documents are not required but information such as name/location of manufacturer, license number for manufacturer, classification of license are required to be written in application document.</p>

Items	Taiwan	Japan
5. Requirement of Technical Standard for Approval	The requirements of technical standard are almost the same such as Mechanical Test, Compatibility between Fixture and Abutment and etc.	
6. Clinical investigation	Clinical investigation is required if there is no predicted products.	
7. Biological Evaluation of Medical Devices	The Biological Evaluation is almost the same besides the Stimulation test.	
	Not mentioned	“使用模擬試驗(Stimulation test)” is required.

Items	Taiwan	Japan									
8. Others	For products less than 3.25 mm in diameter or less than 7 mm in length, the intended use shall be limited to "maxillary central/lateral incisors".	<p>For products out of scope in the chart below, additional evaluation will be required.</p> <table border="1" data-bbox="1025 479 1798 743"> <thead> <tr> <th>Size</th> <th>One-piece</th> <th>Two-piece</th> </tr> </thead> <tbody> <tr> <td>Maximum Diameter</td> <td>3.0-6.0mm</td> <td>3.0-7.0mm</td> </tr> <tr> <td>Length</td> <td>13.5-23.8mm</td> <td>6.0-22.0mm</td> </tr> </tbody> </table> <p>note: two-piece products less than 3.8mm in diameter and less than 6.25mm in length are excepted from the chart</p>	Size	One-piece	Two-piece	Maximum Diameter	3.0-6.0mm	3.0-7.0mm	Length	13.5-23.8mm	6.0-22.0mm
Size	One-piece	Two-piece									
Maximum Diameter	3.0-6.0mm	3.0-7.0mm									
Length	13.5-23.8mm	6.0-22.0mm									



# Overview


1. Product Registration(PR) WG
2. Regulations on Product Pre-market Registration in Taiwan
3. Regulations on Product Pre-market Registration in Japan
4. Review Requirements of Dental Implants in Taiwan and Japan
- 5. Case Study**
6. Summary



# Case Study Introduction (1/2)

藥求安全 食在安心


- Case study item: Dental Implant
- Working process:



# Case Study Introduction (2/2)

藥求安全 食在安心

- Working Process (*cont.*):



# Guidance for Pre-clinical Testing of Endosseous Implant in Taiwan (1/2)

藥求安全 食在安心

- **Scope:**

## F.3640 Endosseous Implant

An endosseous dental implant is a device made of a material such as titanium or titanium alloy, that is intended to be surgically placed in the bone of the upper or lower jaw arches to provide support for prosthetic devices, such as artificial teeth, in order to restore a patient's chewing function.

# Guidance for Pre-clinical Testing of Endosseous Implant in Taiwan (2/2)

藥求安全 食在安心

- **Product description and specification**

Including **Product spec., Drawing, Structure, Material, Chemical Composition, Surface modification** and etc..

- **Safety and performance data**

1. Biocompatibility tests, including **Cytotoxicity, Sensitization, Irritation/Intracutaneous reactivity, Acute systemic toxicity, Subacute and subchronic toxicity, Genotoxicity** and **Implantation. Chronic toxicity** and **Carcinogenicity** are only for new material.
2. Performance tests, including **Mechanical test, Compatibility,** and **Corrosion test.**



# Approval Standard for Dental Implant in Japan (1/3)

藥求安全 食在安心

- Issued in 2009 (QA was also issued in 2012)
- **Scope**
  - Same materials, substantially equivalent surface processing, substantially equivalent surface condition
  - Immediate loading type and short term loading type are exempt
  - Required to meet certain technical criteria
  - Products whose structure, procedure for use, efficacy are significantly different are exempt

If the product is out of scope of this approval standard, the applicant needs to make additional explanation with necessary evaluation.



# Approval Standard for Dental Implant in Japan (2/3)

藥求安全 食在安心

## Main technical criteria

- **Structure**

Form, Size, Assemblage

- **Material**

- **Physical property**

1. Surface processing (Details of processing, roughness after processing)
2. Fatigue test
3. Bending test\*
4. Bending elasticity test\*

\*: These tests are applied for products made from ceramics or polymer material



# Approval Standard for Dental Implant in Japan (3/3)

藥求安全 食在安心

## Main technical criteria

- **Chemical property**

1. Resolvability or resolvability\*;
2. Water absorbability\*;
3. Corrosion resistance\*\*

- **Bio-compatibility**

- **Risk Analysis** including residues due to surface processing and fatigue strength

\*: These criteria are applied for products made from ceramics or polymer material

\*\* : This criterion is applied for products made from metal whose surface processing is not equivalent to the existing product.





# Case Introduction (1/2)

藥求安全 食在安心

Items	
<b>Instruction for use</b>	<ol style="list-style-type: none"><li>1. Made from processed pure grade IV titanium.</li><li>2. Comprised of dental implants and abutments.</li><li>3. Indicated for surgical placement in the upper and lower jaw arches, to provide a root form means for single or multiple unit prosthetic appliance attachment to restore a patient's chewing function.</li><li>4. Designed for conventional two-stage and single stage procedures for single and multiple unit prosthetics</li><li>5. Used for immediate loading procedure</li><li>6. R-ray irradiation</li></ol>

# Case Introduction (2/2)

藥求安全 食在安心

## Items

### Specification

Product Name	Dental Implant System	Project Number	RD0001	Page	4/22
Document Number	BI-IM-B001	Revise date	2012.08.01	Version	D
Item	Specification characteristic	Specification Instruction			Remark
2	Total Length of Implant(mm)	Bone type: 8.0±0.1、10.0±0.1、12.0±0.1、14.0±0.1 Tissue type: 9.8±0.1、11.8±0.1、13.8±0.1、15.8±0.1			
3	Outer Diameter of Root(mm)	3.3±0.02、4.0±0.02、4.7±0.02			
4	Platform Legth(mm)	Bone type: 0.3+0.1、0.4+0.1 Tissuse type: 1.8+0.1			
5	Diameter of neck (mm)	3.3±0.02、4.1±0.02、4.8±0.02、5.5±0.02			
6	Type (Shape) Characteristic	Such as: internal hexagonal feature, flat axial surface features on implants, fins, threads, or vertical anti-rotati slots.			

# Review Comparison (1/3)



藥求安全 食在安心

Items	Taiwan	Japan
-------	--------	-------

## Material and structures

Did not mention about the surface modification process

1. Are bone-type and tissue-type developed with the same concept of development?
2. The concept of development of each type showing details of structure (especially for screw), size, surface processing should be explained.



向更多資訊 <http://>

# Review Comparison (2/3)

藥求安全 食在安心

Items	Taiwan	Japan
<b>Biocompatibility tests</b>	<p>No findings.</p> <p>Notes:</p> <ol style="list-style-type: none"><li>1. Biocompatibility tests meet the requirements of ISO 10993 and “Guidance for Pre-clinical Testing of Endosseous Implant”.</li><li>2. Chronic toxicity and Carcinogenicity are not necessary due to the pure grade IV titanium.</li></ol>	<p>No findings.</p>



# Review Comparison (3/3)

Items	Taiwan	Japan
<b>Functional tests</b>	<ol style="list-style-type: none"><li>1. The fatigue test should be performed with full parts.</li><li>2. Should evaluate whether the test sample is the worst case of the fatigue test.</li><li>3. Should evaluate whether the Shear force and Corrosion test be performed.</li></ol> <p>Note: Compatibility test can be waived because the manufacturer claimed that the dental implant is compatible with their abutment.</p>	<ol style="list-style-type: none"><li>1. Fatigue test should be conducted with specimen which is assembled with components.</li><li>2. Applicant should explain which components are assembled for the test and why it is thought as a weakest case.</li><li>3. Applicants should explain what kind of technology enable the product be used for immediate loading procedure.</li></ol>

# Overview

1. Product Registration(PR) WG
2. Regulations on Product Pre-market Registration in Taiwan
3. Regulations on Product Pre-market Registration in Japan
4. Review Requirements of Dental Implants in Taiwan and Japan
5. Case Study
- 6. Summary**



# Summary

藥求安全 食在安心

- In 2014, PR working group put effort into comparison of the review requirements of dental implants and hemodialysis.
- PR working group implemented a case study program of dental implant to understand the review differences between Taiwan (TFDA) and Japan (PMDA) in 2016.
- For the review findings, there are no significant differences between TFDA and PMDA. Only few dissimilar findings existed in “Material and Structure” and “Functional tests” sections.

**Thank you for your attention!**

