

1 **Biphasic Isophane Insulin Human (Genetical Re-**
2 **combination) Injectable Aqueous Suspension**

3 二相性イソフェンインスリン ヒト(遺伝子組換え)水性懸
4 濁注射液

5
6 Biphasic Isophane Insulin Human (Genetical Recombi-
7 nation) Injectable Aqueous Suspension is an aqueous sus-
8 pension for injection.

9 It contains not less than 95.0% and not more than 105.0%
10 of the labeled Insulin Unit of insulin human (genetical re-
11 combination) (C₂₅₇H₃₈₃N₆₅O₇₇S₆: 5807.57).

12 **Method of preparation** Prepare as directed under Injections,
13 with Insulin Human (Genetical Recombination) Injection and Iso-
14 phane Insulin Human (Genetical Recombination) Injectable
15 Aqueous Suspension.

16 **Description** Biphasic Isophane Insulin (Genetical Recombina-
17 tion) Injectable Aqueous Suspension is a white aqueous suspen-
18 sion. When allowed to stand, it separates into a white precipitate
19 and colorless supernatant liquid, and the precipitate returns to the
20 suspension state on gentle shaking.

21 When it is examined microscopically, the precipitate mostly
22 consist of fine, oblong crystals of 1 to 30 μm in major axis, and
23 does not contain amorphous substances or large aggregates.

24 **Identification** Adjust Biphasic Isophane Insulin (Genetical Re-
25 combination) Injectable Aqueous Suspension to pH between 2.5
26 and 3.0 with dilute hydrochloric acid: the precipitate dissolves,
27 and the solution is clear and colorless.

28 **pH** Being specified separately when the drug is granted ap-
29 proval based on the Law.

30 **Purity (1)** Desamido substance—Perform the test with 20 μL
31 of the sample solution obtained in the Assay as directed under Liq-
32 uid Chromatography <2.01> according to the following conditions.
33 Determine each peak area by the automatic integration method,
34 and calculate the amount of them by the area percentage method:
35 the amount of the peak, having the relative retention time of about
36 1.3 to human insulin, is not more than 1.5%.

37 *Operating conditions*—

38 Proceed as directed in the operating conditions in the Assay.

39 *System suitability*—

40 System performance: Proceed as directed in the system
41 suitability in the Assay.

42 Test for required detectability: Pipet 1 mL of the sample
43 solution, add 0.01 mol/L hydrochloric acid TS to make exactly 50
44 mL. Confirm that the peak area of human insulin obtained with 20
45 μL of this solution is equivalent to 1.4 to 2.6% of that obtained
46 with 20 μL of the sample solution.

47 System repeatability: Dissolve Insulin Human RS in 0.01 mol/L
48 hydrochloric acid TS so that each mL contains about 4 Insulin
49 Units. When the test is repeated 6 times with 20 μL of this solution
50 under the above operating conditions, the relative standard
51 deviation of the peak area of human insulin is not more than 2.0%.

52 (2) High-molecular mass protein—Shake gently, take a suitable
53 volume of Biphasic Isophane Insulin (Genetical Recombina-
54 tion) Injectable Aqueous Suspension, add 4 μL of 6 mol/L hydro-
55 chloric acid TS for every mL of the solution, and mix until the
56 solution becomes clear. Perform the test with 100 μL of this solu-
57 tion as directed under Liquid Chromatography <2.01> according to
58 the following conditions. Determine each peak area by the auto-
59 matic integration method, and calculate the amount of them by the
60 area percentage method: the total amount of the peaks other than
61 human insulin is not more than 2.0%.

62 *Operating conditions*—

63 Detector, column temperature, mobile phase and flow rate:
64 Proceed as directed in the operating conditions in the Purity (2)
65 under Insulin Human (Genetical Recombination).

66 Column: A stainless steel column 7.8 mm in inside diameter and
67 30 cm in length, packed with hydrophilic silica gel for liquid
68 chromatography.

69 Time span of measurement: From the retention time
70 corresponding to the exclusion volume of the size-exclusion
71 column to the completion of the elution of human insulin.

72 *System suitability*—

73 System performance: Proceed as directed in the system
74 suitability in the Purity (2) under Insulin Human (Genetical
75 Recombination).

76 Test for required detectability: Pipet 1 mL of the sample
77 solution, and add 0.01 mol/L hydrochloric acid TS to make exactly
78 50 mL. Confirm that the peak area of human insulin obtained with
79 100 μL of this solution is equivalent to 1.4 to 2.6% of that obtained
80 with 100 μL of the sample solution.

81 **Extractable volume** <6.05> It meets the requirement.

82 **Foreign insoluble matter** <6.06> Perform the test according to
83 Method 1: it meets the requirement.

84 **Sterility** <4.06> Perform the test according to the Membrane fil-
85 tration method: it meets the requirement.

86 **Soluble Insulin Human** Being specified separately when the
87 drug is granted approval based on the Law.

88 **Zinc content** Shake gently, pipet a volume of Biphasic Isophane
89 Insulin (Genetical Recombination) Injectable Aqueous Suspen-
90 sion, equivalent to 300 Insulin Units, and add 0.01 mol/L hydro-
91 chloric acid TS to make exactly 50 mL. If necessary, further add
92 0.01 mol/L hydrochloric acid TS to make exactly 100 mL, and use
93 this solution as the sample solution. Separately, pipet a suitable
94 volume of Standard Zinc Solution for Atomic Absorption Spec-
95 troscopy, dilute with 0.01 mol/L hydrochloric acid TS to make
96 three solutions containing 0.20 μg , 0.60 μg and 1.20 μg of zinc
97 (Zn: 65.38) in each mL, respectively, and use these solutions as
98 the standard solutions. Perform the test with the sample solution
99 and standard solutions as directed under Atomic Absorption Spec-
100 troscopy <2.23> according to the following conditions, using the
101 0.01 mol/L hydrochloric acid TS as the blank, and calculate the
102 amount of zinc (Zn: 65.38) in the sample solution by using the

103 calibration curve obtained from the absorbances of the standard
 104 solutions: 10 – 40 μg per 100 Insulin Units.
 105 Gas: Combustible gas – Acetylene.
 106 Supporting gas – Air.
 107 Lamp: Zinc hollow cathode lamp.
 108 Wavelength: 213.9 nm.

109 **Assay** Shake gently, pipet 10 mL of Biphasic Isophane Insulin
 110 (Genetical Recombination) Injectable Aqueous Suspension, and
 111 add exactly 40 μL of 6 mol/L hydrochloric acid TS. Pipet 2 mL of
 112 this solution, add 0.01 mol/L hydrochloric acid TS to make exactly
 113 5 mL, and use this solution as the sample solution. Separately,
 114 weigh accurately a suitable amount of Insulin Human RS, dissolve
 115 exactly in 0.01 mol/L hydrochloric acid TS to make a solution so
 116 that each mL contains about 40 Insulin Units, and use this solution
 117 as the standard solution. Perform the test with exactly 20 μL each
 118 of the sample solution and standard solution as directed under Liq-
 119 uid Chromatography <2.01> according to the following conditions.
 120 Determine the peak areas, A_{TI} and A_{SI} , of human insulin and those,
 121 A_{TD} and A_{SD} , of the desamido substance having the relative reten-
 122 tion time of about 1.3 to human insulin, respectively, in each solu-
 123 tion.

124 Amount (Insulin Unit) of human insulin ($\text{C}_{257}\text{H}_{383}\text{N}_{65}\text{O}_{77}\text{S}_6$) in 1
 125 mL

$$126 = (M_S \times F) / D \times (A_{\text{TI}} + A_{\text{TD}}) / (A_{\text{SI}} + A_{\text{SD}}) \times 1.004$$

$$127 \times 5 / 2$$

128 M_S : Amount (mg) of Insulin Human RS taken

129 F : Labeled Unit (Insulin Unit /mg) of Insulin Human RS

130 D : Volume (mL) of 0.01 mol/L hydrochloric acid TS used to
 131 dissolve Insulin Human RS

132 *Operating conditions* –

133 Detector: An ultraviolet absorption photometer (wavelength:
 134 214 nm).

135 Column: A stainless steel column 4.6 mm in inside diameter and
 136 15 cm in length, packed with octadecylsilanized silica gel for
 137 liquid chromatography (5 μm in particle diameter).

138 Column temperature: A constant temperature of about 40°C.

139 Mobile phase: A mixture of phosphoric acid-sodium sulfate
 140 buffer solution (pH 2.3) and acetonitrile for liquid
 141 chromatography (3:1). Adjust the mixing ratio of the component
 142 of the mobile phase so that the retention time of human insulin is
 143 between 10 and 17 minutes.

144 Flow rate: 1.0 mL per minute.

145 *System Suitability* –

146 System performance: When the procedure is run with 20 μL of
 147 human insulin desamido substance-containing TS under the above
 148 operating conditions, human insulin and human insulin desamido
 149 substance are eluted in this order with the resolution between these
 150 peaks being not less than 2.0, and the symmetry factor of the peak
 151 of human insulin is not more than 1.8.

152 System repeatability: When the test is repeated 6 times with 20
 153 μL of the standard solution under the above operating conditions,

154 the relative standard deviation of the peak area of human insulin
 155 is not more than 1.6%.

156 **Containers and storage** Containers – Hermetic containers.

157 Storage – Light-resistant, at a temperature between 2°C and
 158 8°C avoiding freezing.

159