

Particle size analysis by dynamic light scattering

The method is based on ISO standard 22412:2017 *Particle Size Analysis - Dynamic light scattering (DLS)*.

1 INTRODUCTION

Dynamic light scattering (DLS) can be used to measure the average particle size and the broadness of the size distribution of submicron particles dispersed in a liquid.

Average particle size and particle size distribution are important characteristics of dispersed systems such as emulsions, suspensions and liposome formulations.

DLS can be used to measure particles in the submicron range and is therefore particularly suitable for the particle size analysis of dispersed systems that are composed of randomly moving particles measuring up to approximately 1 μm .

2 PRINCIPLE

Submicron particles dispersed in a liquid, and that are free from sedimentation, are subject to a perpetual random movement, known as Brownian motion. When these particles are irradiated with a laser, scattered light intensity from the moving particles fluctuates depending on their diffusion coefficients. The intensity of the scattered light from larger particles fluctuates more slowly, because larger particles move more slowly and conversely the intensity of the scattered light from smaller particles fluctuates more rapidly.

In dynamic light scattering measurements the diffusion dependent fluctuations of the scattered light intensity are measured and analyzed. The diffusion coefficient and the particle diameter are related by the Stokes-Einstein equation.

$$x = \frac{kT}{3\pi\eta D} \times 10^{12}$$

x : hydrodynamic diameter of an equivalent spherical particle (nm)

k : Boltzmann constant ($1.38 \times 10^{-23} \text{ J} \cdot \text{K}^{-1}$)

T : absolute temperature (K)

η : viscosity of the dispersing medium ($\text{mPa} \cdot \text{s}$)

D : translational diffusion coefficient ($\text{m}^2 \cdot \text{s}^{-1}$)

The intensity of light scattered by the diffusing particles exhibits a time dependency which can be described as either a time-dependent phase shift or as a spectral frequency shift.

Based on these concepts, the time-dependent intensity of the scattered light is processed either by photon correlation spectroscopy (PCS) or by frequency analysis.

In PCS, the time-dependent intensity of the scattered light is correlated with a time-delayed copy of itself (autocorrelation function) or with the signal from a

46 second detector (cross-correlation function). Both the auto- and cross-correlation
47 function of a monodisperse system decay exponentially with correlation time.
48 The decay rate depends on the fluctuation of the scattered light as a function of
49 particle size (slower for large particles and faster for small particles).

50 In frequency analysis, the frequency-based power spectrum of the scattered
51 light is analysed. For a monodisperse system, the power spectrum is a
52 Lorentzian type function.

53 These two methods are mathematically equivalent. The time-based
54 autocorrelation function in PCS is equal to the Fourier transform of the
55 frequency-based power spectrum in frequency analysis. Therefore, the average
56 diameter (\bar{x}_{DLS}) and the polydispersity index (PI), which indicates the broadness
57 of the particle size distribution, can be evaluated with each method.

58 Different mathematical approaches are applied for data evaluation, including
59 a Laplace inversion for particle size distribution or the cumulants method to
60 evaluate the time-based autocorrelation function.

61 Two types of optical detection are used with DLS instruments: homodyne
62 detection, in which only the scattered light is measured and heterodyne
63 detection, in which the scattered light and a portion of the incident light are
64 combined for interference.

65

66 **3 INSTRUMENT**

67 The measuring system typically consists of:

- 68 (i) A laser: a monochromatic and coherent laser beam polarized with its electric field
69 component perpendicular to the plane formed by the incident light beam and
70 light-receiving optical axes (vertical polarization), illuminating the sampler in the
71 measuring cell.
- 72 (ii) A sample holder: the sample holder must maintain the temperature of the sample to
73 within ± 0.3 °C.
- 74 (iii) Optics and a detector: a light detector positioned at a fixed angle relative to the
75 incident laser beam measuring (usually at only one scattering angle) the apparent
76 scattered light intensity (i.e. the sum of the scattered light from all the particles in the
77 scattering volume) at appropriate intervals. When a polarization analyser is included, it
78 is positioned so that the transmittance of the vertically polarized light is maximized.
- 79 (iv) A correlator (photon correlation spectroscopy) or spectrum analyser (frequency
80 analysis).
- 81 (v) A computation unit and data processing software (some computation units also
82 function as correlators or spectrum analysers).

83

84 **4 INSTRUMENT PERFORMANCE**

85 As the particle sizes obtained by DLS are not relative values calculated using
86 standard particles but absolute values based on the first principle, calibration
87 cannot be performed.

88 However, the performance of the instrument must be checked after it is first
89 installed or if abnormal performance is suspected using particles with a certified
90 diameter; it is recommended to repeat this check at least once a year thereafter.
91 Certified reference materials with values assigned for DLS must be used.

92 Dispersions of polystyrene latex with narrow size distribution with certified
93 particle diameter of about 100 nm or other suitable size can be used.

94 The measured average particle size must be within the stated range of the
95 certified reference material expanded by 2% on each side. The polydispersity
96 index must be smaller than 0.1 and the relative standard deviation of at least
97 five repeated measurements on a sample must be equal or lower than 2 %.

98 99 **5 PROCEDURE**

100 **5.1 SAMPLE PREPARATION**

- 101 (i) Samples consist of the test substance well-dispersed in a liquid. To eliminate the
102 influence of multiple light scattering, their concentration must be within an
103 appropriate range. The dispersion medium must:
- 104 a. be non-absorbing at the wavelength of the laser;
 - 105 b. be compatible with the materials used in the instrument;
 - 106 c. not induce particle dissolution, swelling or agglomeration/aggregation;
 - 107 d. have a known refractive index different from that of the test substance;
 - 108 e. have known viscosity to within ± 2 per cent over the operational range of
109 temperature to be used;
 - 110 f. be clean and free of dust for low background scattering.
- 111 (ii) The range of the particle concentration is determined so that the results of the
112 measurements do not vary significantly. The range is determined beforehand based on
113 measurements of systematically diluted samples.

114 It is also important to remove dust since it may affect the measurement, and
115 to prevent its re-introduction during preparation. Typically, scattered light
116 signals from the dispersion medium used for sample dilution must be
117 undetectable or very weak. If large fluctuations in the scattered light signals
118 accompanied by abnormally strong signals are recorded or if light spots appear
119 in the path of the laser light in the sample, foreign particles are likely to have
120 been incorporated into the sample. In such cases, further purification of the
121 dispersion medium is necessary (by filtration, distillation, etc.) before use. The
122 lower limit of the particle concentration range is determined mainly so that
123 scattered light from the dispersion medium and foreign particles will not affect
124 the measurement.

125 When water is chosen as the dispersion medium, use of fresh distilled water or
126 desalted and filtered (nominal pore size 0.2 μm) water is recommended.

127 Long-range electrostatic interactions arising between highly charged particles
128 may affect the measurement result. In such cases, a small amount of salt (for
129 example, about 10^{-2} mol/L sodium chloride) may be added to the dispersion
130 medium to reduce the effect. Air bubbles may also appear in the test sample,
131 particularly when measuring a refrigerated sample at room temperature, and
132 are to be avoided.

133 **5.2 TEST PROCEDURE**

134 Switch the instrument on and allow it to warm up.

135 Clean the measurement cell if necessary. The degree of cell washing
136 required depends on the conditions of the measurement. When an individually
137 packaged clean disposable cell is used, cleaning is not necessary. When a cell is
138 intended to be washed, it is rinsed with water or an organic solvent. If required,
139 a non-abrasive detergent may be used.

140 Place the measurement cell containing the sample in the sample holder, and
141 wait until temperature equilibrium is reached between the sample and the

142 sample holder. It is recommended to measure and maintain the temperature to
143 within $\pm 0.3^{\circ}\text{C}$.

144 Perform a preliminary measurement of the sample, and set the particle
145 concentration within the appropriate range (see Sample preparation).

146 Perform the measurement with the appropriate measuring time and
147 number of acquisitions.

148 Record the average particle diameter and the PI for each measurement. If
149 the measured values are dependent on the particle concentration, take the
150 extrapolated infinite dilution values of the average particle diameter and the
151 PI (or the measured values at the lowest particle concentration).

152 Confirm that no significant settling has occurred in the sample at the end of
153 the measurement. The presence of a sediment indicates that the sample may
154 have agglomerated/aggregated or precipitated, or that it may not be a suitable
155 candidate for DLS.

156 **5.3 REPEATABILITY**

157 The achievable repeatability of the method mainly depends on the
158 characteristics of the test substance (emulsion/suspension; robust/fragile;
159 broadness of its size distribution; etc.), whereas the required repeatability
160 depends on the purpose of the measurement. Mandatory limits cannot be
161 specified in this chapter, as repeatability (different sample preparations) may
162 vary appreciably from one substance to another. However, it is good practice to
163 aim for repeatability at a relative standard deviation of not more than 10 per
164 cent [$n \geq 3$] for \bar{x}_{DLS} .

165

166 **6 RESULTS**

167 The test report must include the average particle diameter and PI.

168 It must state the dispersion medium used, the refractive index, viscosity and
169 temperature of the test sample, and give sufficient information about the
170 measurement system, including the principle of measurement (PCS or frequency
171 analysis), optical configuration (homodyne or heterodyne) and observation angle.
172 The measuring time or number of acquisitions, the sample (nature,
173 concentration and preparation method), the dispersion conditions, and the
174 measurement cell type must also be described. As the results depend on the
175 instrument, data analysis program and optical model used, these details must
176 also be provided.

177

178 **7 GLOSSARY**

179 (i) Average particle diameter, \bar{x}_{DLS} : Harmonic intensity-weighted averaged
180 particle diameter expressed in nanometres.

181 (ii) Polydispersity index, PI: dimensionless measure of the broadness of the
182 particle size distribution.

183 (iii) Scattering volume: section of the incident laser beam viewed by the
184 detector optics. Its order of magnitude is typically 10^{-12} m^3 .

185 (iv) Scattered intensity, count rate: intensity of the light scattered by the
186 particles in the scattering volume as measured by a detector. In PCS, the
187 number of photon pulses per unit time expressed in counts per second. In
188 frequency analysis, the photodetector current which is proportional to the

- 189 scattered light intensity.
- 190 (v) Viscosity, η : viscosity of the dispersion medium in $\text{mPa} \cdot \text{s}$.
- 191 (vi) Refractive index, n : dimensionless refractive index of the dispersion
- 192 medium at the wavelength of the laser.