

1 Installation Environment, Basic Handling 2 Method, and Precautions for Weighing of 3 a Balance <G1-8-182>

4 (はかり(天秤)の設置環境, 基本的な取扱い方法と秤量
5 時の留意点 <G1-8-182>)

6
7 The mass to be weighed should be usually larger than the
8 minimum weight. Before weighing, tools used for weighing
9 should be prepared and kept tidy and in order (cleaning), and
10 the sensitivity of a balance should be adjusted. The installa-
11 tion environment of a balance, basic handling method, and
12 precautions for weighing are shown as follows.

13 1. Installation environment of a balance

14 It is desirable to install a balance at a place whose sur-
15 rounding environment is unchanged at any time in a room not
16 too spacious, avoiding vibration sources, ventilation ports,
17 and wall surfaces that receive heat radiation from the room
18 light and direct sunlight. It is desirable to use a balance in the
19 corner of a room or near a large pillar where the impact of
20 vibrations is considered to be small. It is also desirable that a
21 weighing table (anti-vibration table, vibration isolated table,
22 etc.) on which a balance is installed has enough mass by itself,
23 it is so rigid that it is not distorted vertically even if the load
24 of a heavy object is applied, and magnetism and electrifica-
25 tion are taken into consideration. Especially for a balance
26 with the number of readable digits of not more than 0.1 mg,
27 precautions must be taken for installation or relocation be-
28 cause displayed values become unstable from the transmis-
29 sion of minor vibrations that cannot be sensed by humans to
30 a sample itself or the reaction of the weighing sensor of the
31 balance to minor vibrations. For the aspect of maintenance
32 and control, an environment without sudden temperature
33 changes, which may cause condensation, is required to avoid
34 the deterioration of the components of a balance. The instal-
35 lation environment for a balance as an electronic device
36 should be within a temperature range from 5°C to 40°C and
37 relative humidity range from 20% to 80%, and a relative hu-
38 midity of not less than 45% is desirable when considering the
39 effect of static electricity.

40 2. Operation check before the use of a balance

41 The following items should be checked before using a bal-
42 ance.

43 2.1. Securement of preheating standby time

44 After turning on the power, secure a preheating standby
45 time to stabilize the internal temperature of the detector. It is
46 desirable to secure the following preheating standby time of
47 not less than 30 minutes when the number of readable digits
48 is 10 mg or more, not less than 1 hour when it is 1 mg to 10
49 mg, not less than 2 hours when it is not more than 1 mg, and
50 not less than half a day when it is not more than 0.01 mg.

51 2.2. Check of installation condition

52 Check that a balance is level using the level gauge installed,
53 for example, the air bubble in the gauge is positioned at the
54 center. For the level adjustment, it is desirable to visually
55 check whether the balance is unstable and whether there is no
56 gap between the legs of the balance, which contact the weigh-
57 ing table, and the surface of the table.

58 2.3. Implementation of sensitivity adjustment

59 In the case of a balance equipped with a sensitivity adjust-
60 ment function (equipped with an internal weight for adjust-
61 ment), it is possible to perform appropriate sensitivity adjust-
62 ments depending on the surrounding temperature condition
63 at the zero point and around the balance's capacity. The
64 higher the resolution, the greater the effect of changes in sen-
65 sitivity, and measurement errors due to changes in sensitivity
66 generally become relatively large from the zero point to
67 around the mass weighed. As for a device that is not equipped
68 with a sensitivity adjustment function, it is desirable to man-
69 ually adjust sensitivity using the weight of around the bal-
70 ance's capacity.

71 3. Cleaning

72 To avoid weighing an object other than the target object,
73 cleaning should be performed periodically. If the structure of
74 a balance is understood, and the balance can be simply dis-
75 mantled and cleaned, clean each part using a glass cleaner
76 and lint free cloth frequently and keep the weighing pan and
77 the inside of the weighing room clean.

78 4. Elimination of external factors that affect weighing 79 results

80 External factors that affect weighing results must be elim-
81 inated as much as possible. In the case of samples that are
82 prone to moisture absorption, adsorption, volatilization, or
83 evaporation, measures must be implemented according to the
84 characteristics of the sample to prevent the deviation of
85 weighed values. For example, when a hygroscopic sample is
86 weighed, weighing with good repeatability is possible if a
87 balance is installed in a constant temperature and humidity
88 box, and the sample is weighed after it is acclimated to the
89 specific temperature and humidity conditions in advance. Ex-
90 ternal factors that affect weighing results other than the char-
91 acteristics of a sample itself are shown below.

92 4.1. Temperature differences between around the 93 weighing pan and a sample (including sampling con- 94 tainer)

95 Temperature differences between around the weighing pan
96 and a sample occur because of the refrigerated storage of a
97 sample, bringing a sample from outside of the room with a
98 different temperature, heat treatment, heat conduction by
99 body temperature, and other factors. If the temperature of a
100 sample and sampling container is higher than the temperature
101 in the weighing room, slight upward wind (convection)

102 generates around the weighing pan, and the sample and sam-
103 pling container are pushed up by the phenomenon, resulting
104 in a decrease or instability of displayed values. If the relation-
105 ship of the temperature is contrary, an opposite trend appears.
106 These phenomena are physical phenomena that occur around
107 the weighing pan and therefore cannot be avoided even if a
108 balance is equipped with a windshield device. Therefore,
109 weighing should be performed under the condition where the
110 temperature in the weighing room of a balance and the tem-
111 peratures of a sample and a sampling container are as equiv-
112 alent as possible.

113 **4.2. Wind due to air conditioners, etc.**

114 When the weighing pan is affected by an airflow generated
115 by an air conditioner, access to the weighing room by people,
116 and the weighing operation of a person who performs meas-
117 urement, the displayed values become unstable. To suppress
118 the impact of such wind, a windshield device should be in-
119 stalled to prevent direct wind onto the weighing pan. Or, re-
120 locate it to a place where there is no wind. When a balance
121 equipped with an open/close door is used under the condition
122 where it is affected by direct wind, it is important not to open
123 the open/close door wider than necessary.

124 **4.3. Static electricity**

125 When using a sample such as powders and a sampling con-
126 tainer, which are likely to be electrified by friction, or the in-
127 side of a weighing room is in a low humidity condition of
128 relative humidity of 40% or lower, weighing results were af-
129 fected by displayed values that fluctuate upward or down-
130 ward due to the action of the force of electrical charges. The
131 following items are considered to prevent such static electric-
132 ity: keep humidity in the weighing room at 45% or higher,
133 wait for dissipation of accumulated static electricity, and
134 change the sampling container to an antistatic container. If
135 these measures cannot be implemented, it is recommended to
136 perform measurement after performing discharging as much
137 as possible using an instrument, such as an ionizer, which
138 neutralizes charged electrical charges or promotes the dissi-
139 pation of the charges. However, avoid using an instrument
140 that directly blows wind, which makes displayed values un-
141 stable, to the weighing pan during discharging.

142

143